



art & science of
amazing protection

NITRILE



POWDER FREE
EXAMINATION
GLOVES

*Quality Crafted,
Trusted **Protection***

Series No.:
20570 Blue Nitrile Powder Free Examination Gloves



BENEFITS OF NITRILE POWDER FREE GLOVES

Trusted by Healthcare: The perfect formulation for healthcare that strikes the correct balance between comfort and protection. Used by millions of healthcare workers around the world. You know you can trust ASAP with your protection.

Compliance and Quality Assurance: Meets the highest industry standards, including Medical Device Regulation (EU) 2017/745 and PPE Regulation (EU) 2016/425, ensuring consistent quality and performance.

Effortless Ambidexterity: With a design that fits both hands seamlessly, our gloves enhance convenience and streamline the fitting process.

Vibrant Blue Colour: Easily identifiable in any setting, making them ideal for food processing and beyond.

Powder-Free Design: Eliminates the risk of powder-related allergies and contamination, ensuring a safer and cleaner experience.

Enhanced Grip: Strategically textured fingers provide superior grip, reducing hand strain and enhancing dexterity during tasks.

Ideal Weight & Comfort: Strikes the perfect balance between tactile sensitivity, comfort and protection.

Versatile Applications: Suitable for various sectors including healthcare, dental, automotive, cleaning, retail, and more.

Regulations

- Medical Device Regulation (EU) 2017/745
- PPE Regulation (EU) 2016/425
- Food Contact Regulation (EU) 2020/1245 of Regulation (EU) No 10/2011
- REACH Regulation

Harmonized Standards

- EN ISO 21420:2020
- EN 374-1:2016+A1:2018
- EN 374-4:2019
- EN 374-5:2016
- EN 455-1:2020
- EN 455-2:2015
- EN 455-3:2015
- EN 455-4:2019

Quality Assurance

- ISO 9001:2015
- ISO 13485:2016
- ISO 14001:2015



At ASAP, we are committed to hygiene control and quality assurance. Proper hygiene standard is practiced throughout the development of all ASAP products from raw materials handling, processing, production, inspection, to our finished product to deliver high quality products while limiting risk of cross-contamination.

Look for the Hygiene Matters™ logo, quality and hygiene you can trust.





Colour Option:



Healthcare



Food & Beverage



Cleaning



Series No.:
20570 Blue Nitrile Powder Free Examination Gloves

Series Size Codes

Extra Small	Small	Medium	Large	Extra Large
XS, 6	S, 7	M, 8	L, 9	XL, 10
20571	20572	20573	20574	20575

Product Specifications

Design	Ambidextrous, Finger Textured Surface, Beaded Cuff
Colour	Blue
Acceptance Quality Level (AQL)	1.5
Packing Mode	100 pcs per box, 10 boxes per carton

Dimension Specifications

Glove Size	Palm Width (mm)	Length (mm)		Thickness Single Wall (mm)	
		EN 455	Cuff (25±5 from bead)	Palm (centre of palm)	Finger (13±3 from tip)
XS, 6	75 ± 5	Min. 240	0.06 ± 0.02	0.07 ± 0.02	0.10 ± 0.02
S, 7	85 ± 5				
M, 8	95 ± 5				
L, 9	105 ± 5				
XL, 10	115 ± 5				

Physical Properties Specifications

	EN 455 Force at Break (N)
Before Aging	Min. 6.0
After Aging	Min. 6.0

Packaging Dimensions

Inner	220 x 110 x 65mm
Carton	340 x 230 x 230mm

Powder Residue

Powder Free (mg/glove)	Max. 2
------------------------	--------

Instructions For Use

Description - Nitrile Powder Free Examination Gloves, Non-sterile, Single Use Only.

Intended Use - ASAP nitrile glove is a disposable glove product worn to protect the hand of wearer against mechanical action whose effects are superficial, cleaning materials of weak action and easily reversible effects.

How To Don Gloves - Inspect the gloves to ensure there are no pinholes or tears. If gloves are ambidextrous, they can be worn on either hand. If not, align the glove's fingers and thumb with the proper hand before donning. Insert five fingers into the cuff and pull the cuff over the wrist. Check for a secure fit around the fingers and palm. The cuff should fit snugly around the wrist.

How To Doff Gloves - After use, users should visually check the glove and remove any contamination from the outer surface before removing the gloves from the hands. Grasp the outside edge of the glove near the wrist. Peel the glove away from the hand, turning it inside out. Hold it in the opposite glove hand. Slide an ungloved finger under the wrist of the remaining glove, be careful not to touch the outside of the glove. Peel the remaining glove off from the inside, creating a "bag" containing both gloves. Discard.

Disposal - Properly dispose of all used nitrile glove. Follow your institution's policies for use and disposal of these gloves.

Storage - Store in a dry place. Avoid excessive heat (30°C). Exposed product should be shielded from direct sunlight, intense artificial light, x-ray machines, and other source of ozone.

Shelf Life - Three years from the manufacturing date.

Warning - These gloves are for single and transient use only.

Caution - This information does not reflect the actual duration of protection in the workplace and the differentiation between mixtures and pure chemicals. The chemical resistance has been assessed under laboratory conditions from samples taken from the palm only (except in cases where the glove is equal or over 400mm - where the cuff is also tested) and relates only to the chemical tested.

It can be different if the chemical is used in a mixture. It is recommended to check that the gloves are suitable for the intended use because the conditions at the workplace may differ from the type test depending on the temperature, abrasion, and degradation. When used, protective gloves may provide less resistance to the dangerous chemical due to changes in the physical properties. Movements, snagging, rubbing, degradation caused by the chemical contact etc. may reduce the actual use time significantly.

For corrosive chemicals, degradation can be the most important factor to consider in selection of chemical resistant gloves. This product contains nitrile rubber, which may cause allergic reactions in individuals who are known or suspected to be allergic to nitrile rubber. If an allergic reaction occurs, stop using immediately and consult a physician. This product is not made of natural rubber latex.

EN ISO 374

Chemical Permeation (EN ISO 374-1:2016+A1:2018/Type C)	Level	Mean Degradation % (EN ISO 374-4:2019)	Degradation levels indicate the change in Puncture Resistance of the glove after exposure to the challenge chemical.
K 40% Sodium Hydroxide	5	-17.7	
P 30% Hydrogen Peroxide	0	12.9	
T 37% Formaldehyde	4	-22.7	

EN ISO 374

EN 16523-1:2015+A1:2018 Classification of Permeation Performance Level

Measured Breakthrough Time (min)	>10	> 30	> 60	> 120	> 240	> 480
Permeation Performance Level	1	2	3	4	5	6

The penetration levels have been assessed under laboratory conditions and relates only to the tested specimen.

Resistance against Bacteria and Fungi - PASS
Resistance against Virus - PASS

ASAP INTERNATIONAL SDN BHD
No. 1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor,
Malaysia.

T : +603 5191 0166
F : +603 5191 0702
E : info@whyasap.com
W : www.whyasap.com

ASAP INNOVATIONS LTD.
Unit 7, The Courtyard,
Fonthill Business Park,
Fonthill Road, Dublin,
D22 XA07, Ireland.

T : +353 1466 1660
E : info@whyasap.ie
W : www.whyasap.ie

ASAP INNOVATIONS (UK) LTD.
13, Diamond Court,
Opal Drive, Fox Milne,
Milton Keynes,
MK15 0DU, United Kingdom.

T : +44 (0) 1908 732700
E : info@whyasap.co.uk
W : www.whyasap.co.uk

